



A pesquisa que constrói o futuro

Japan-Brazil Cooperation
Improvements

Paulo
DLF

- The most urgent improvement is in the system for determining the capacity of the EUT at 50% of its nominal capacity.
- Other improvements suggested by training participants:
 - Measurement of air flow velocity at the temperature sensors, during the entire stabilization and test period (under analysis)
 - Perform a spreadsheet review, as a non-compliance was detected (ISO 17025): some spreadsheet cells were accidentally not protected

OBRIGADO

santos@cepel.br

(21) 2598-6297



A pesquisa que constrói o futuro



A pesquisa que constrói o futuro

MINISTÉRIO DE
MINAS E ENERGIA

